

ECOLAB HEALTHCARE ANZ www.ecolab.com AU 1300 425 529 www.healthcare-au.ecolab.com NZ 0800 425 529 www.healthcare-au.ecolab.com

# **OptiPro Multi-Enzyme**

Concentrated multi-enzymatic manual detergent concentrate for manual and ultrasonic cleaning

#### ARTG 235775



### **Features**

- A concentrated multi-enzymatic detergent with outstanding cleaning performance for manual and ultrasonic cleaning of surgical instruments and manual cleaning of endoscopes. Provides improved removal of hard-to-clean soils, reducing soaking time and brushing.
- Optipro is a concentrated multi-enzymatic detergent with outstanding cleaning performance for manual and ultrasonic cleaning of surgical instruments and manual cleaning of endoscopes.

Outstanding Performance

- Advanced, multi-enzymatic technology reduces the soaking and scrubbing needed to remove surgical soils, improving process efficiencies and user experience.
- Versatile and multi-purpose can be used as a pre-soak, manual cleaner, ultrasonic detergent, and flexible endoscope detergent in all water and soil conditions.
- Low-foaming, crystal clear use solution allows visibility of instrumentation in the sink.
- Neutral pH, non-corrosive and free-rinsing formulation is compatible with surgical instruments and flexible endoscopes.
- Effectively removes hard-to-see soil residues that can remain after cleaning.

#### Advanced Technology

- Utilizes the latest enzyme technology specific for surgical soils.
- Optimized surfactants for removal of lipids.
- Suitable for use in a wide range of water temperatures.

#### Sustainable Solution

- 2X concentrated liquid provides more uses per litre.
- Phosphate-free to help reduce environmental impact.
- Contains biodegradable surfactants.

Product Code	Product Description	Pack Size	Quote
71006911	OptiPro Multi-Enzyme Manual Concentrate Deterget, 3.78 L	4 Pk	



ECOLAB HEALTHCARE ANZ www.ecolab.com AU 1300 425 529 www.healthcare-au.ecolab.com NZ 0800 425 529 www.healthcare-au.ecolab.com

## Safety Information

Do not mix with any other products. For further information, safety instructions and health warnings, see product label and Safety Data Sheet. ALWAYS READ THE LABEL. FOLLOW DIRECTIONS OF USE.

## **Directions For Use**

### **Manual Cleaning of Flexible Endoscopes:**

- The cleaning solution should be made fresh for every scope.
- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Proceed with the flexible endoscope cleaning process per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

#### **Ultrasonic Cleaning:**

- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4mL/L of warm water.
- Use solution temperature should not exceed 60°C.
- Complete a degas cycle prior to cleaning.
- Run appropriate ultrasonic cleaning cycle per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

### **Automated Flushing of Flexible Endoscopes:**

- The cleaning solution should be made fresh for every scope.
- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Proceed with the automated flushing of flexible endoscope per the device manufacturer's written IFU.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

### **Pre-Soak and Manual Cleaning of Medical Instruments:**

- Dilute OptiPro Multi-Enzymatic Manual Detergent Concentrate at 1-4 mL/L of warm water.
- Use solution temperature should not exceed 43°C.
- Pre-soak instruments as needed until soil is dissolved and removed from the instruments. Soak for a minimum of 1-5 minutes.
- Additional soak time or light mechanical cleaning with an instrument brush may be necessary depending on instrument complexity or age of soil. Refer to IFU for brushing and cleaning.
- Rinse and flush items per the device manufacturer's written IFU, then proceed with the decontamination process.

## **Relevant Product**

- Soluscope Sprint
- <u>Soluscope Sprint PT</u>
- EndoSelect Cleaning Brushes

- <u>Bandelin Sonorex</u>
- <u>Bandelin Sonomic</u>
- Bandelin Trison



ECOLAB HEALTHCARE ANZ www.ecolab.com AU 1300 425 529 www.healthcare-au.ecolab.com NZ 0800 425 529 www.healthcare-au.ecolab.com

• Belimed UW565 Ultrasonic

Bandelin Sonoboard

# **Applications**

- Infection Control
- Healthcare Facilities
- Central Sterile Departments
- Endoscopy
- Residential Aged Care

### **Resources**

- safety data sheet (sds)
- Video
- Brochure